



OptoSelect[®] 1750b Chip and OptoSelect[®] 1750b Chip – Research

PRODUCT INSERT

Purpose

The OptoSelect[®] 1750b chip (750-00018) is intended for use on the Beacon[®] and Beacon Select[™] optofluidic systems. The OptoSelect[®] 1750b chip – Research (750-08334) is intended for use only on specially configured Beacon and Beacon Quest[™] optofluidic systems. These nanofluidic chips enable cell import, culture, assay, and export on the Beacon systems. The chips are disposable and single-use only. They include Wetting Solution, which is a proprietary solution used to prepare chips for cell loading.

Kit Contents

Kit 750-00018, OptoSelect 1750b chip for the Beacon[®] and Beacon Select[™] optofluidic systems

Components	Unit Quantity	Volume/ Unit	Uses/ Unit	Storage Conditions	Part Number
1750b Chip	1 chip	1	1	Room temperature	750-00018
Wetting Solution ^a	2 mL	1	1	4°C	

a. Ships with the chip, but can be ordered separately (520-00009).

Kit Contents

Kit 750-08334, OptoSelect 1750b chip – Research for specially configured Beacon and Beacon Quest[™] systems


Components	Unit Quantity	Volume/ Unit	Uses/ Unit	Storage Conditions	Part Number
1750b Chip – Research	1 chip	1	1	Room temperature	750-08334
Wetting Solution ^a	2 mL	1	1	4°C	

a. Ships with the chip, but can be ordered separately (520-00009).

FOR RESEARCH USE ONLY. Not for use in diagnostic procedures.

Chip Specifications


Feature	Specification
Number of NanoPen® chambers	1,758
Chip volume (channel + pens)	6.8 µL
Number of channels	4
Channel width	225 µm
Chip height	40 µm
NanoPen volume	1.70 nL
NanoPen width	130 µm
NanoPen length	370 µm



OptoSelect Chip



Single NanoPen Chamber



Chip Map

Instructions for Use

Instructions for use and handling can be found in the following documents on the Support Portal:

- *Beacon® Optofluidic Systems User Manual* (MAN-000003)
- *Opto® CLD 1.0 Workflow User Manual* (MAN-08036)
- *Opto® CLD 2.0 Workflow User Manual* (MAN-08096)

Chip Handling, Cleaning, and Disposal

Chips should only be handled using powder-free nitrile gloves and should be held at the corners.

Clean the chips before installing them into the nest using a pre-soaked 99% isopropyl alcohol (IPA) wipe. Fold the IPA wipe three times and wipe the glass surface of the chip from left to right in one smooth, continuous motion while applying moderate downward pressure. Dispose of the wipe after use. Using a new pre-soaked 99% IPA wipe, wipe the surface of the ferrules and dispose of the wipe. Once the IPA has evaporated from the surface of the chip and the ferrules, close the nest lid.

Dispose of chips in a biohazardous waste container according to local, state, and federal regulations.

Safety

Use standard laboratory safety protocols and always wear appropriate personal protective equipment (PPE). Read and understand the safety data sheets (SDSs) before handling chemicals. To obtain SDSs, email support@brukercellularanalysis.com.

Expiration

Wetting Solution and OptoSelect chips are shipped with a minimum of 6 months available shelf life. The expiration date for Wetting Solution is located on the individual vial. The expiration date for the OptoSelect chip is located on the outer chip package label. The Wetting Solution and OptoSelect chips are not under warranty past their expiration dates.

Legal

The OptoSelect 1750b chip and OptoSelect 1750b chip – Research product insert and its contents are proprietary to Bruker Corporation. This product insert is intended to be used, and should only be used, in connection with the use of the Beacon optofluidic systems. No other use is permitted without the prior written consent by an authorized officer of Bruker. No right, title, or interest in any Bruker intellectual property rights are conveyed or granted, expressly or by implication, by this product insert. Review the documents listed in [Instructions for Use](#) PRIOR to using the OptoSelect 1750b chips.

Bruker reserves the right to change the information in this product insert without notice.

Bruker Cellular Analysis
brukercellularanalysis.com
+1 (510) 858-2855
sales@brukercellularanalysis.com

Technical Support
support.brukercellularanalysis.com
+1 (888) 254-5595 ext. 4
support@brukercellularanalysis.com

FOR RESEARCH USE ONLY. Not for use in diagnostic procedures.

Bruker Cellular Analysis, Beacon, Beacon Select, Beacon Quest, NanoPen, Opto, OptoSelect, and the Bruker logo are trademarks of Bruker Corporation. All other marks are the property of their respective owners.